

VELACUR: INITIAL RESULTS IN DISCRIMINATION OF PATIENTS AND HEALTHY VOLUNTEERS BASED ON FIBROSIS AND STEATOSIS

Michael Curry¹, Edward Tam², Caitlin Schneider³, Noha Abdelgelil⁴, Tarak Hassanien⁴, Nezam Afdhal¹

¹ Beth Israel Deaconess Medical Center, Boston, MA

² Vancouver, BC, CA

³ Sonic Incytes, Vancouver, BC


⁴ Southern California Background Research Center, Coronado

INTRODUCTION

Diagnostic core liver biopsy for assessing liver health has significant drawbacks including sampling bias and patient pain, leading to poor patient compliance¹. Liver elasticity, or stiffness, is correlated with histological liver fibrosis stage and offers a non-invasive method for assessment and monitoring of liver fibrosis². Ultrasound tissue attenuation is correlated with liver fat content, and is an increasingly important parameter given the rise of non-alcoholic fatty liver disease. Elasticity is most commonly measured by shear wave speed in Transient Elastography (FibroScan®)³ or shear wavelength in steady-state elastography as in Magnetic Resonance Elastography (MRE)⁴. MRE is the non-invasive imaging method that provides the most accurate assessment of clinical fibrosis stage, when compared to biopsy⁵.

AIM

To evaluate the ability of a Velacur™ prototype to discriminate between healthy volunteers and those with clinically diagnosed non-alcoholic fatty liver disease (NAFLD) or previous Hepatitis C virus (HCV) infection. Exploratory objectives looked at the concordance and correlation between Velacur and MRI.




METHODS

We use a prototype elasticity measurement system (Velacur™, previously known as Liver Incytes, Sonic Incytes, BC, Canada) comprising of an ultrasound probe and an activation unit, to excite multi-frequency (40-70 Hz), steady-state shear waves in the patient. As with MRE, the shear wave field is measured over a volume and used to produce the average volumetric spatial elasticity.


Data processing algorithms are based on previous work^{6,7} and data presented here include new algorithms for automatic shear wave data quality assessment and for automatic vessel detection within the liver.

The multi-frequency approach provides an opportunity for averaging to reduce measurement uncertainty. Attenuation measurements are made from the ultrasound data captured simultaneously with elasticity measurements. **Both patients (n = 86) and healthy volunteers (n = 50) participated in this study and were scanned with both Velacur, and FibroScan. A subset of patients agreed to undergo MRE (n = 31).**

RESULTS



Above: The components of the Velacur prototype, including the ultrasound probe, the laptop to run software, the control unit to coordinate the signals and the activation unit which creates shear waves in the patient. A curvilinear abdominal ultrasound probe is used to image the liver volume. Using a sweep motion of the probe, multiple planes of ultrasound data at a depth of 15 cm are acquired over 30 degrees, in the elevational direction. **RIGHT:** A rendering of the general positioning of the probe, liver and measured volume. The blue cone (middle) shows how much of the liver can be seen in the ultrasound image. The cube within the liver shows the large region which is used to measure the average elasticity and attenuation.



Above: The cohort separation results for FibroScan and the Velacur prototype. Velacur was able to differentiate the two cohorts with similar AUC to FibroScan. Only healthy volunteers without steatosis or patients with steatosis were included in the steatosis separation.

Characteristic	Value
Total Enrolled	N = 136
Healthy	N = 50
NASH/NAFLD	N = 59
HCV SVR	N = 27
Age (mean (range))	49 (21,77) years
BMI (mean (range))	27.9 (17, 42) kg/m ²
AST (mean ± stdev)	42 ± 35 U/L
ALT (mean ± stdev)	39 ± 44 U/L
Platelets (mean ± stdev)	206 ± 88 10 ⁹ /L
FibroScan Probe	M = 87 XL = 49

CONCLUSIONS

The results from this study demonstrate the promise of this technique for quantitative non-invasive assessment of fibrosis and attenuation in volunteers and patients with chronic liver disease. The ability for this Velacur™ prototype, to discriminate between healthy volunteers and patients with liver disease is comparable to FibroScan®, used in current clinical practice. All AUROCs were greater than 0.85 for both fibrosis and attenuation measurements.

There is excellent concordance between Velacur elasticity and MRE, at 0.80. The correlation between Velacur attenuation and MRI-PDFF results is also excellent.

The Velacur prototype has proved that it:

- Can recognize advanced disease
- Can scan patients with high BMI, up to 42 kg/m²
- Has larger measured volume than current clinical care
- Shows higher correlation with MRE
- Correlates well with MRI-PDFF
- Has benefits of portability and accessibility

These qualities make it suitable for point-of-care diagnosis and regular patient monitoring during and after treatment.

ACKNOWLEDGEMENTS


The study sponsors would like to acknowledge the help and participation of all the volunteers and patients who agreed to take part in this study, as well as the work and dedication of the staff at all the clinical sites.

REFERENCES

[1] Wong, R., Ahmed A. Complications of liver biopsy. *Gastrointestinal Emergencies*. 2016 Jan 20:77.
 [2] Hall, T., et al. RSNA/QIBA: Shear wave speed as a biomarker for liver fibrosis staging. *IEEE IUS*, pp. 397-400, 2013.
 [3] Catheline, J.-L. et al. "Diffraction field of a low frequency vibrator in soft tissues using transient elastography." *IEEE Trans. Sonics Ultrason.*, 46(4), pp. 1013-1019, 1999.
 [4] Manduca, A., et al. "Magnetic resonance elastography: Non-invasive mapping of tissue elasticity." *Med. Image Anal.*, 5(4), pp. 237-254, 2001.
 [5] Singh, S., et al. "Diagnostic performance of magnetic resonance elastography in staging liver fibrosis: a systematic review and meta-analysis of individual participant data." *Clinical Gastro. and Hepatology* 13.3 (2015): 440-451.
 [6] Abeysekera, J., et al. Vibro-elastography: Absolute elasticity from motorized 3d ultrasound measurements of harmonic motion vectors. In *IEEE Intl. Ultrasonics Symposium*, pp. 1-4, 2015.
 [7] Baghani, A., et al. Real-time quantitative elasticity imaging of deep tissue using free-hand conventional ultrasound. *Intl. Conference on Medical Image Computing and Computer-Assisted Intervention*, pp. 617-624, 2012.

CONTACT INFORMATION

mcurry@bidmc.harvard.edu
caitlin@sonicincytes.com



Far Above: Graphs of the concordance coefficient (ρ_c) between Velacur and MRE in black and FibroScan and MRE in red. Dotted line shows the 1 to 1 line, while the solid lines are linear fits which cross the origin.

Near Above: Graphs of the correlation coefficient (r^2) between Velacur ACE and MRI-PDFF in black and FibroScan CAP and MRI-PDFF in red. These graphs include some healthy volunteers from a previous pilot study, and 3 patients were removed as invalid scans due to user or technical error.